

INQUA®

Equipments Under Test (EUT):

INQUA® Inhalator

REF: BR021000, BR025000

INQUA® Inhalator 1000

REF: BR030000

belAir® Inhalator

REF: BR040000

**Electromagnetic compatibility – Guidance and manufacturer's
declaration IEC 60601-1-2: Edition 4.0; 2014-02**

1 ESSENTIAL PERFORMANCE

There are no essential performance characteristics according to the risk assessment.

2 ELECTROMAGNETIC ENVIRONMENT

The device¹ is intended for use in the electromagnetic environment specified below. The customer or the user of the device¹ should assure that it is used in such environment.

Emission test

Emission test	Compliance	Electromagnetic environment – Guidance
RF emissions CISPR 11	Group 1	The device ¹ uses RF energy only for its internal function. Therefore, its RF-emission is very low and not likely to cause any interference nearby electronic equipment.
	Class B	The device ¹ is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Passed	

Immunity test

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – Guidance
Electrostatic discharge (ESD) IEC 61000-4-2	Contact: ±8 kV Air: ±15 kV	Contact: ±8 kV Air: ±15 kV	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient / burst IEC 61000-4-4	5/50 ns, 100 kHz, ±2 kV	5/50 ns, 100 kHz, ±2 kV	Mains power quality should be similar to that of a typical commercial or hospital environment.

1) Device under test: INQUA® Inhalator (REF: BR021000, BR025000), INQUA® Inhalator 1000 (REF: BR030000) and belAir® Inhalator (REF: BR040000).

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – Guidance
Surge IEC 61000-4-5	1.2/50 (8/20) μ s LtL: ± 1 kV LtG: ± 2 kV	1.2/50 (8/20) μ s LtL: ± 1 kV LtG: ± 2 kV	Mains power quality should be similar to that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	0 % U_T for 0.5 cycle (1 phase) 0 % U_T for 1 cycle 70 % U_T for 25/30 cycles (50/60 Hz) 0 % U_T for 250/300 cycles (50/60 Hz)	0 % U_T for 0.5 cycle (1 phase) 0 % U_T for 1 cycle 70 % U_T for 25/30 cycles (50/60 Hz) 0 % U_T for 250/300 cycles (50/60 Hz)	Mains power quality should be similar to that of a typical commercial or hospital environment. When the user of the device ¹ requires continuous function in the event of disruption of supply, it is recommended the device ¹ is operated from an uninterruptible power supply or battery.
Power frequency (50 Hz / 60 Hz) magnetic field IEC 61000-4-8	30 A/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Conducted RF IEC 61000-4-6 Amplitude modulated	150 kHz – 80 MHz 3 V ISM and amateur radio bands 6 V <input checked="" type="checkbox"/> 80 % / 1 kHz	150 kHz – 80 MHz 3 V ISM and amateur radio bands 6 V <input checked="" type="checkbox"/> 80 % / 1 kHz	WARNING: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the device ¹ , including cables, specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – Guidance
Radio-frequency electromagnetic field Amplitude modulated	IEC 61000-4-3	80 MHz – 2.7 GHz 10 V/m <input checked="" type="checkbox"/> Home Healthcare <input type="checkbox"/> Prof. Healthcare <input checked="" type="checkbox"/> 80 % / 1 kHz	WARNING: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the device ¹ , including cables, specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.
Proximity fields from RF wireless communications equipment	IEC 61000-4-3	380 – 390 MHz 27 V/m; PM 50 %; 18 Hz 430 – 470 MHz 28 V/m; (FM ±5 kHz, 1 kHz sine) PM; 18 Hz 704 – 787 MHz 9 V/m; PM 50 %; 217 Hz 800 – 960 MHz 28 V/m; PM 50 %; 18 Hz 1700 – 1990 MHz 28 V/m; PM 50 %; 217 Hz 2400 – 2570 MHz 28 V/m; PM 50 %; 217 Hz 5100 – 5800 MHz 9 V/m; PM 50 %; 217 Hz	WARNING: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the device ¹ , including cables, specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

INQUA®



INQUA GmbH

Moosdorfstraße 1

82229 Seefeld • GERMANY

Tel.: +49 (0) 8152 99 34-0

Fax: +49 (0) 8152 99 34-20

info@inqua.de • www.inqua.com